

**Office of the Quality Assurance Manager  
Quality Control Laboratory  
Department of Fisheries, Chattogram.**


**Proficiency Test Information  
(01.01.2018 to 30.06.2019)**


Name of the QC Laboratory	Name and Address of PT Provider	Period of PT	Matrix	Name of The Test Parameters	Results/ Z-Score	Comments
QC Laboratory, Chattogram.	TESTVERITAS, Italy	Round of September, 2018	Fish & Shrimp	Tetracyclines	Satisfactory	ELISA Screening
			Fish & Shrimp	Chloramphenicol	Satisfactory	
			Fish	Histamine	Satisfactory	
		Round of November, 2018	Fish	Malachite Green Leucomalachite Green	Satisfactory	
	BIPEA, France	October, 2018	Shellfish Mix	Hg	0.35	Analysed by ICPMS
				Cd	0.62	
				As	0.67	
				Cr	1.09	
				Pb	4.50	
		April, 2019	Mollusk Mix	Hg	2.81	
				Cd	1.89	
				As	-0.69	
				Cr	0.24	
		June, 2019	Salmon	Pb	3.32	
				Hg	0.35	
				Cd	0.45	
				As	1.43	
	National Food Agency, Sweden	January, 2019	Food	Standard Plate Count (SPC)	-0.317	Analyzed by ISO Method.
					-0.054	
					-0.338	
National Food Agency, Sweden.	January, 2019	Food	<i>Salmonella spp.</i>	Satisfactory	Analyzed by ISO Method.	


A

/

QC Laboratory , Chattogram.	National Food Agency, Sweden.	January, 2019	Food	<i>Listeria monocytogenes</i>	Satisfactory	Analysed by ISO Method.
	National Food Agency, Sweden.	January, 2019	Food	Pathogenic <i>Vibrio spp.</i>	Satisfactory	Analysed by ISO Method.
	Aquaculture Pathology Lab, University of Arizona, USA.	February, 2019	Shrimp	<i>White Spot Syndrome Virus (WSSV)</i>	Satisfactory	Analysed by Nested PCR Technique
	Aquaculture Pathology Lab, University of Arizona, USA.	February, 2019	Shrimp	Infectious Myonecrosis Virus (IMNV)	Satisfactory	Analysed by Nested PCR Technique
	Aquaculture Pathology Lab, University of Arizona, USA.	February, 2019	Shrimp	<i>AHND (EMS)</i>	Satisfactory	Analysed by Nested PCR Technique
	Aquaculture Pathology Lab, University of Arizona, USA.	February, 2019	Shrimp	<i>IHHNV</i>	Satisfactory	Analysed by Nested PCR Technique
	Aquaculture Pathology Lab, University of Arizona, USA.	February, 2019	Shrimp	<i>NHP-B</i>	Satisfactory	Analysed by Nested PCR Technique.

  
 04.08.19  
**Md. Mozanur Rahman**  
 Bsc Fisheries (Hons) MS BCS(Fisheries)  
 Fish Inspector & Quality Control Officer  
 Quality Control Laboratory  
 Department of Fisheries  
 Chattogram

  
 04.08.19  
**Salma Begum**  
 Bsc Fisheries (Hons). MS. BCS(Fisheries)  
 Fish Inspection & Quality Control Officer  
 Quality Control Laboratory  
 Department of Fisheries  
 Chattogram

  
 04.08.19  
**(Md. Abdur Razzaque)**  
 Quality Assurance Manager  
 Quality Control Laboratory  
 Department of Fisheries, Chattogram.